

Colon Hydrotherapy Inspection Checklist- Final

Name of the Facility: _____

Date of Inspection: ____/____/____

Ref.	Description	Yes	No	N/A	Remarks
4.	Facility Requirements				
4.2	Colon hydrotherapy may be added to an outpatient care facility that provides at least one of the following specialties: gastroenterology, internal medicine or general surgery				
4.6	Specialized plumbing is required for the human waste released during the treatment				
4.7	Sediment filter in addition to an activated carbon filter attached to the hot and cold water connection of the treatment room is available				
4.8	Second filter system attached to the colon hydrotherapy device itself in accordance with the manufacturer specification is available				
4.10	Installation of an ultraviolet sterilizer at the main supply specifically for the closed pressure systems				
4.12	Cleansing solution recommended by the manufacturer should be used				
5.	The procedure/ treatment room				
5.1.	Consultation room & procedure/treatment rooms must be separate & their size abides by the outpatient care regulation				
5.2.	Floors must be fully washable. Carpet floor are not allowed.				
5.3.	Room must have its own toilet & washing facilities				
5.4.	Room must have an extraction system				
5.5.	Water supply of drinking water is available on the premises				
5.6.	Separate hands free basin with hot & cold running water supplied through a single outlet with liquid soap & paper towels				
5.7.	Have a toilet & shower for client's exclusive use located in the procedure room or as an ensuite				
5.8.	Have a paper towel on the client couch				

5.10.	Facility must provide clean & comfortable facilities for the patient/ client to change				
5.11.	Clean gowns, robes & towels should be provided				
5.13	Crash cart trolley must be available in the procedure room				
6.	Procedure Requirements				
6.1	Patient care:				
6.1.9.	Under no circumstances should the colonic irrigation equipment be connected directly to a drinking water supply system				
6.3.	Water temperature:				
6.3.1	-Minimum allowed temperature is 31°C -Maximum allowed temperature is 39°C -Average treatments temperature of 34°C to 37°C				
7.	Infection control				
7.1.	Hygiene:				
7.1.1.	There must be disinfectant in the room at all times e.g. sterilizing alcohol or Milton solution				
8.	Disposables				
8.1.	It is universally recommended to use only disposable kits.				

Inspectors: (Name & Signature)

1. _____
2. _____
3. _____
4. _____